

Levitating Laser Heating Techniques for
Synchrotron Radiation Experiments.

A Thesis submitted to Aberystwyth University
in the subject of Physics for the Degree of
Master of Philosophy.

By

Stephen David Fearn

February 2009

Abstract

This project is about the design, development and construction of individual components and equipment associated with a 125 watt Carbon Dioxide laser heated aerodynamic furnace that was capable of levitating and melting a 2mm sphere in a dedicated containerless nozzle. The technique allows a specimen to be heated to 3000° Kelvin, and the purity of the sample to be investigated without contamination. This novel experiment was designed and tested at Aberystwyth University, and subsequently installed and operated on the Synchrotron Radiation Source 6.2 x-ray beamline. Experiments using the laser heated aerodynamic furnace on refractory oxides at high temperatures have allowed the successful investigation of structural dynamics at the nano and atomic level, by the use of Rapid2 detectors, with particular interest in the recalescence and supercooled regions, with subsequent investigations resulting in the discovery and observations of the polyamorphic rotor.

Acknowledgements

I would like to take this opportunity for thanking all those people who contributed both directly and indirectly, to the completion of this thesis.

Prof. G. Neville Greaves for showing the confidence in allowing me to do this thesis and in giving me a free reign in the design and development of this exciting experiment.

Dr Tudor Jenkins for his guidance and support in the writing of this thesis.

Dr Dave Langstaff for his help and support in the initial development of the LabVIEW software.

Dr Martin Wilding, Dr Florian Kagl and all the academic staff of the Institute of Mathematics and Physics.

Dr Louis Hennett for his guidance of the experiment as it has developed over the last few years.

Mr Clive Willson for his continual support.

Mr John Parry and Mr David Francis for their outstanding contribution in the mechanical fabrication of the experiment, and their continual support in the development of the instrumentation.

Mr Les Dean and Mr Dave Lewis for the help in electronic designs.

Dr Chris Martin 6.2 beamline Daresbury Laboratories for all the outstanding contributions towards the success of this facility, from initial installation to operational experiments.

Mr Alfie Nield for all his technical help and support during the initial installations onto the Daresbury Beamline.

All the staff of the Daresbury laboratory.

Levitating Laser Heating Techniques for Synchrotron Radiation Experiments

Contents

Chapter 1	Overall Design	5
	1.1. Bench	7
	1.2. Optics table	8
	1.3. Breadboard	8
	1.4. Polycarbonate box	9
	1.5. Pyrometer mounts	10
	1.6. Sample holder	11
	1.7. Mirrors	11
	1.8. External Interfacing	12
	1.9. References	14
	1.10. Appendix Detail plans of beamline 6.2	
Chapter 2	Carbon Dioxide Laser	15
	2.1. Inductive heating	15
	2.2. Laser Operation	17
	2.3. Synrad evolution laser	18
	2.4. RF3000 laser power supply	20
	2.5. Environmental control facilities	21
	2.6. UC2000 Laser controller	23
	2.7. Laser Diode	25
	2.8. Auxiliary Signal Connector	26
	2.9. Conclusions	29
	2.10. References	29

Chapter 3	Optical Pyrometry	30
	3.1 Choice of Optical Pyrometry	31
	3.2 Black body radiation	31
	3.3 Luxtron optical pyrometer	32
	3.4 Basic pyrometer operation	33
	3.5 Pyrometer Input Options	34
	3.6 Sample Alignment	36
	3.7 Daresbury Synchrotron Experimentation	37
	3.8 LabVIEW	38
	3.9 Conclusions	39
	3.10 References	40
	3.11 Luxtron calibration report	41
Chapter 4	Mass Flow Controller	42
	4.1 Volume reservoir	42
	4.2 Brooks mass flow controller	43
	4.3 Description of mass flow controller	44
	4.4 Computer use of DDE Software	45
	4.5 Conclusions and Future experiments	47
	Appendix	
	MODEL 5850S	48
	Flowtech Calibration Certificate	49
Chapter 5	Laser Power Meter	50
	5.1 Introduction	50
	5.2 Probe	51

	5.3	RS232 serial communication	52
	5.4	Experimental procedure	52
	5.5	Further investigations	53
	Appendix		
		Certificate of Calibration Power Instrument	54
		Certificate of Calibration Powermeter Probe	55
Chapter 6		Optics	56
	6.1	Objectives and initial configuration	56
	6.2	Optics table	57
	6.3	Beam Splitter	58
	6.4	Mirrors	58
	6.5	Focusing lenses	58
	6.6	Fixed focal length mirrors	59
	6.7	Mirror alignment	61
	6.8	Daresbury configuration	62
	6.9	Remote control mirrors	63
	6.10	Thermal mirrors	65
	6.11	Final mirror choice	66
	Appendix		
		Circuit diagram of servo controller	67
		Acknowledgements	67
Chapter 7		Nozzle Design and Sample Stage for liquid levitation	68
	7.1	Boron Nitride nozzles	68
	7.2	Sample sphere characteristics during levitation	71
	7.3	Nozzle dimensions	72
	7.4	Aluminium nozzles	73

	7.5	Water cooled sample stage	74
	7.6	Sample stage vertical adjustment	75
	7.7	Future developments of the sample stage	77
Chapter 8		X-ray beam alignment	
79			
	8.1	X-ray beam characteristics	79
	8.2	Beam size	81
	8.3	Establishing datum Line	81
	8.4	Beam alignment and size	82
	8.5	GDA software	84
	8.6	XAFS table	84
	8.7	Horizontal nozzle alignment	84
	8.8	Vertical nozzle position	86
	8.9	Detector systems	88
	8.10	Conclusions	89
	8.11	References	90
Chapter 9		LabVIEW Software Control	91
	9.1	LabVIEW	92
	9.2	LabVIEW interfacing	93
	9.3	D/A conversion	94
	9.4	LabVIEW software	96
	9.5	Laser and software status	97
	9.6	Manufacturing anomalies and interference	98
	9.7	Optical Pyrometry LabVIEW control software	100
	9.8	Vertical sample stage interface	101
	9.9	External triggering	103

	9.10	Future developments	104
Chapter 10		Laser Shutter	105
	10.1	Laser beam shutter	106
	10.2	Construction and operation	106
	10.3	Shutter Status	109
	10.4	Remote indicator	109
	10.5	Remote shutter key control box	110
	10.6	System upgrade	112
	10.7	Conclusions	112
Chapter 11		Safety Protocol and video diagnostics	113
	11.1	Safety factors for the use of class 4 lasers	113
	11.2	Interlock control system	115
	11.3	Experimental alignment	117
	11.4	Laser Emergencies	119
	11.5	Emergency Kill switches	119
	11.6	Laser safety glasses	120
	11.7	Polycarbonate enclosure	120
	11.8	Smoke detector	121
	11.9	Ceramic beam stops	121
	11.10	Welding glass filters	121
	11.11	Compressed air	122
	11.12	Compressed gases	122
	11.13	Chilled water	122
	11.14	Video camera surveillance	123

11.15	Radio cameras	124
11.16	Microscope camera	125
11.17	Sample plane video camera	126
11.18	External safety camera	126
11.19	High speed camera	127
11.20	Recording video imaging of liquid drops	127

Appendices

1. Photo-Sonics Specification sheet
2. Interlock and shutter circuit diagram
3. Standing orders

Chapter 12	Results	130
12.1	Melting	130
12.2	Recalescence	131
12.3	Supercooling	132
12.4	High and low density transition	133
12.5	Future experiments	138
12.6	Conclusions	139
12.6.1	Twin laser system	139
12.6.2	Twin laser shutter	139
12.6.3	Video monitoring	140
12.7	References	140
12.8	Published papers	141

List of Figures

Figure A. Schematic operation of the laser furnace in conjunction with the equipment on the 6.2 beamline at the Daresbury Synchrotron Radiation Source.	3
Figure B. Molten sample levitating at 2000° K in the x-ray beam, picture taken through Pb window of the 6.2 beamline hutch.	4
1.1 Side elevation schematic, illustrating optimal sample position with respect to x-ray beam position and wide (WAX) and small (SAX) angle detectors.	6
1.2 Schematic plan and side view, illustrating initial component layout with respective dimensions.	7
1.3 Side and front view of all experimental components in position assembled on the optics bench on the dedicated laser stand at the Aberystwyth Material Physics Laboratory.	8
1.4 Plan view looking into the polycarbonate enclosure during furnace development.	9
1.5 Internal view of the polycarbonate box during initial setup on the 6.2 beamline.	10
1.6 Beamline view of the levitating furnace during setup with the guide laser illuminating the levitating sample; also in view are the sample stage cooling pipes and the gas delivery pipes.	11
1.7 Levitation furnace installed into the 6.2 beamline at Daresbury Laboratory.	12
2.1 Inductive heating and levitating techniques.	16
2.2 Schematical laser operation.	17
2.3 Synrad evolution laser attached via adjustable legs to optics bench.	19
2.4 RF3000 controller and dedicated power supply.	20
2.5 a) Chilled water pump unit for regulating temperature through the heat exchanger.	22
b) Heat exchanger for dissipating heat from the laser head through the demineralised water system at Daresbury Laboratory.	22
2.6 UC2000 laser controller incorporating manual and remote control of the laser.	23
2.7 Computer serial communication serial port and LabVIEW digital to analogue communication port.	24

2.8	Externally coupled laser diode used for optical alignment of the carbon dioxide laser.	25
2.9	Remote monitoring indicator box for showing RF3000 and laser status; also included is solenoid activation indicator to confirm that the shutter is operational.	27
3.1	Position of the optical pyrometer head with respect to the high temperature levitating sample.	30
3.2	Graph illustrating relationship between relative energy and wave length.	32
3.3	Luxtron optical pyrometer.	32
3.4	Schematic block diagram of the operative of the Luxtron optical pyrometer.	33
3.5	Initial development layout of the pyrometer position prior to beamline installation.	34
3.6	Final position using H bridge support enabling the pyrometer to straddle the x-ray delivery tube.	35
3.7	Laser diode adaptor used for optical pyrometer alignment onto the levitating sample.	36
3.8	H bridge pyrometer configuration suitable for straddling the x-ray delivery tube.	37
3.9	RS232 pin identification and protocol.	39
4.0	Computer port communication and Luxtron communication facilities.	39
4.1	Schematic layout requirements of control gas delivery for levitation.	42
4.2	Schematic operation of the Brooks mass flow controller, enabling controlled delivery of gas to the sample head for levitation.	43
4.3	Computer communications port and mass flow controller setup.	45
4.4	RS232 interconnections between digital mass flow controller and computer.	46
4.5	Manufacturer schematics of the Brooks mass flow controller.	48
5.1	Molelectron laser power meter.	50
5.2	Molelectron power head.	51
5.3	Schematic experimental layout of power meter.	52
5.4	Powermeter instrument calibration certificate.	54

5.5	Powermeter probe calibration certificate.	55
6.1	Sample stage position on the 6.2 Beamline with x-ray and laser beam paths.	56
6.2	Schematical component layout of aerodynamic laser-heated furnace.	57
6.3	Schematic experimental layout of plain mirrors and carbon dioxide focusing lenses.	59
6.4	Illustration of relative positions of focused mirrors and respective focal lengths.	60
6.5	Video image taken from the microscope camera illustrating levitating molten sample with heated beam position.	61
6.6	Mirror profiles illustrating vertical and horizontal adjustment, allowing accurate positioning of the laser beam onto the sample.	62
6.7	Confined geometry of the levitator laser furnace installed on the 6.2 beamline.	63
6.8	Server motors illustrating vertical mirror position adjustment.	64
6.9	Master and slave control box used for remotely controlling the accurate position of mirrors.	65
6.10	Customized mirror heat sink for dissipating heat generated by laser power including remote vertical pitch adjustment.	66
6.11	Circuit design of servo controlled motors, adapted for mirror motor control.	67
7.1	Examples of boron nitride nozzles showing the dimensions of the sample throat and gas inlet.	69
7.2	Boron nitride nozzle viewed through electron microscope.	70
7.3	Polished surface structure of Boron nitride nozzle.	70
7.4	Nozzle schematic for cup and cone shaped designs.	72
7.5	Illustration of sample instability during levitation.	72
7.6	Nozzle design and cooled sample stage.	73
7.7	Plan view of cooled sample stage incorporating gas inlet and x-ray beam path.	74
7.8	Profile view of cooled sample stage illustrating optical pyrometer and laser beam path.	75

7.9	Vertical sample stage movement incorporating analogue and digital feedback.	76
7.10	Customised levitator head with sample stage cooling designed to minimise air scatter.	77
7.11	Prototype aerodynamic levitator module designed to handle oxidising liquids and sufficiently small to be installed within the 60 cm radius space within the WAXS detector.	78
8.1	Illustrating the position of the sample holder with respect to the X-ray beam position and the WAXS / SAXS detectors.	80
8.2	Profile illustration position of the sample and levitation stage with respect to the X-ray beam path and the WAXS / SAXS detector.	81
8.3	Theodolite image illustrating central data line taken using a mobile phone camera.	82
8.4	Screen shot view of the GDA software illustrating the position and control of all 5 sets of slits.	83
8.5	X-ray sensitive paper positioned behind the sample, indicating beam position and size.	83
8.6	X-ray beam attenuation graph for horizontal sample position.	85
8.7	Polaroid image illustrating the shadow of drill position and sample holder in x-ray beam.	86
8.8	Drill position on top of the sample stage enabling horizontal position of the x-ray beam.	86
8.9	X-ray attenuation as the drill bit passes through the beam in the vertical position.	87
8.10	Block diagram and specification of Rapid2 detector.	88
8.11	WAXS detector template allowing accurate repositioning of the detector in the x-ray beam.	89
9.1	Instrumentation for laser heating and temperature measurement controlled by LabVIEW.	91
9.2	Flow diagram of LabVIEW software initialisation for computer laser control.	92
9.3	Adlink Technology digital to analogue interface board which gives proportional output to the UC2000 laser controller.	94
9.4	Schematic of digital to analogue convertor used to control UC2000 laser controller.	95

9.5	Analogue output connections of the Adlink Technologies interface board used to control the UC2000 laser controller.	95
9.6	LabVIEW construction block diagram window designed to communicate with the UC2000 laser controller and Luxtron optical pyrometer.	96
9.7	Front panel interface for LabVIEW control of laser heating and temperature monitoring.	97
9.8	Synrad software specification anomaly.	98
9.9	Co plot of pyrometer output and laser power for a levitating super cooled drop, showing intermediate interference spikes.	99
9.10	LabVIEW software flow diagram for initialising and communicating with the Luxtron optical pyrometer.	101
9.11	Detailed vertical adjustment of liquid drop height at different temperatures.	102
9.12	LabVIEW stepper motor control of vertical sample stage, with analogue and digital feedback potentiometers for sample position.	102
10.1	Schematical layout of laser shutter position.	105
10.2	Laser beam shutter schematics.	106
10.3	Internal layout of the laser shutter illustrating the default position of the laser beam stop, the default return springs, interlock beam micro switches, the two way operating pneumatic ram and solenoid actuator.	107
10.4	External layout of laser shutter incorporating solenoid status indicator, pneumatic ram and two way actuator.	108
10.5	Laser shutter circuit diagram, incorporating air pressure switch, beam block position monitoring and status indicators.	109
10.6	Remote laser status indicator panel and solenoid activation indicator.	109
10.7	Remote shutter key control panel, with shutter status indicators, remote shutter closing and emergency laser disable panic button.	111

10.8	Remote shutter control circuit diagram.	111
10.9	Final operational laser shutter with compressed air indicator for monitoring air flow into two-way pneumatic ram.	112
11.1	Schematic diagram of laser levitator furnace interlock system and how it integrates with experimental beam line interlock system.	114
11.2	Experimental hutch interlock hazard display panel on MPW beam line 6.2.	115
11.3	Laser interlock control panel.	117
11.4	Laser emergency stop and shutter control box.	119
11.5	Polycarbonate box on test bench and polycarbonate box installed onto beamline.	120
11.6	Plan layouts of video cameras.	123
11.7	Radio camera on optical mount.	124
11.8	Video footage of liquid sample.	124
11.9	Microscope camera.	125
11.10	Video monitor illustrating the calibrated position of the liquid sample in the x-ray beam.	126
11.11	Photo-Sonic high speed camera remotely controlled to study molten samples.	127
11.12	Experimental control desk incorporating LabVIEW laser and pyrometer control computer, mass flow control computer, and video camera monitoring station and recording facilities.	128
12.1	Real time video capture of levitating molten sample, illustrating laser beam hot spot.	131
12.2	Co plot illustrating heating and cooling cycle of alumina, highlighting the melting stage and recalescence regions as special regions of interest.	132
12.3	SAXS data for supercooled AY15, AY20 and AY25 liquids revealing the intensity of random fluctuation measured at the Synchrotron Radiation Source beamline 6.2.	134
12.4	Changes to the principle peak for AY20 between 2392K and 1515K, measured at the Synchrotron Radiation Source beam line 6.2.	135
12.5	Oscillating rotor.	136

12.6	Sequential rotating images of the liquid sample in the supercooled region.	136
12.7	Video images A to F illustrating horizontal rotation. Fluctuating image brightness for one pixel area compared with pyrometer output for one 4-s cycle. 180° rotations take 600ms, after which the bead is virtually stationary.	137
12.8	Pyrometer readings of oscillating polyamorphic rotor.	138

